

Varicella zoster virus IgA ELISA

Enzyme immunoassay for the qualitative determination of IgA class antibodies against Varicella Zoster virus in human serum or plasma (citrate).

REF**30114080****12x8****2-8°C**

EU:

IVD**CE****IBL INTERNATIONAL GMBH**Flughafenstrasse 52a
D-22335 Hamburg, GermanyPhone: +49 (0)40-53 28 91-0
Fax: +49 (0)40-53 28 91-11IBL@IBL-International.com
www.IBL-International.com

1. INTRODUCTION

Varicella-Zoster Virus (human herpes virus 3, HHV-3) belongs to the a-subfamily of herpesviridae. The virus particles measure about 145 nm in diameter. They consist of doublestranded DNA, are surrounded by an icosahedral protein capsid and an envelope which contains both host cells and viral components. The virus is usually transmitted in respiratory secretions, and a single serotype causes varicella (Chickenpox), a highly infectious childhood disease, and zoster (shingles), a neurodermic disease; both diseases are found worldwide. Varicella is the acute disease which follows primary contact with the virus, whereas zoster is the response of the partially immune host to a reactivation of the varicella virus present in the body in latent form. Varicella is endemic, most commonly affected are children between 2 and 6 years of age. The course of disease is usually mild and complicated only in immunocompromised children. Rare fatal cases show multiple necrotic lesions in brain, lung (varicella pneumonia), kidneys (hemorrhagic nephritis), spleen, bone marrow, and occasionally in the intestinal tract. The lethality of varicella is below 0.1%. In the infrequent adult infections the disease is more severe, and complications are to be expected in about 5% of all cases. Zoster is of low incidence and appears with increasing frequency and severity with advancing age. Usually the process remains localized, generalization is frequently encountered in a state of immunosuppression. Fatal cases are very rare and nearly always caused by an underlying disease.

Species	Disease	Symptoms	Mechanism of Infection
Varicella-Zoster Virus (VZV)	Varicella (primary infection)	Vesicular eruptions of the skin and mucous membranes; exanthema	Transmission by droplet infection-the viruses replicate first in the mucous membranes of the upper respiratory tract and then spread hematogenously. Incubation period: Varicella 14-17 days or zoster probably 7-18 days
	Zoster (secondary infection)	Generally zoster shows more inflammation and destructive changes (i.e. necrosis, hemorrhages)	

The presence of virus resp. infection may be identified by:

- Microscopy: Giemsa stain; electron microscopy; IF
- Serology: Detection of antibody production by ELISA

2. INTENDED USE

The Varicella zoster virus IgA ELISA is intended for the qualitative determination of IgA class antibodies against Varicella-Zoster virus in human serum or plasma (citrate).

3. PRINCIPLE OF THE ASSAY

The qualitative immunoenzymatic determination of IgA-class antibodies against Varicella-Zoster Virus (VZV) is based on the ELISA (Enzyme-linked Immunosorbent Assay) technique.

Microtiter strip wells are precoated with Varicella-Zoster Virus (VZV) antigens to bind corresponding antibodies of the specimen. After washing the wells to remove all unbound sample material horseradish peroxidase (HRP) labelled anti-human IgA conjugate is added. This conjugate binds to the captured Varicella-Zoster Virus (VZV)-specific antibodies. The immune complex formed by the bound conjugate is visualized by adding Tetramethylbenzidine (TMB) substrate which gives a blue reaction product. The intensity of this product is proportional to the amount of Varicella-Zoster Virus (VZV)-specific IgA antibodies in the specimen. Sulphuric acid is added to stop the reaction. This produces a yellow endpoint colour. Absorbance at 450 nm is read using an ELISA microwell plate reader.

4. MATERIALS

4.1. Reagents supplied

- **MTP Varicella-Zoster Virus (VZV) Coated Wells (IgA)**: 12 breakapart 8-well snap-off strips coated with Varicella-Zoster Virus (VZV) antigen; in resealable aluminium foil.
- **SAMPLEDIL IgA Sample Diluent *****: 1 bottle containing 100 mL of buffer for sample dilution; pH 7.2 ± 0.2; coloured yellow; ready to use; white cap.
- **TMB STOP Stop Solution**: 1 bottle containing 15 mL sulphuric acid, 0.2 mol/L; ready to use; red cap.
- **WASHBUF CONC Washing Solution (20x conc.)***: 1 bottle containing 50 mL of a 20-fold concentrated buffer (pH 7.2 ± 0.2) for washing the wells; white cap.
- **ENZCONJ Varicella-Zoster Virus (VZV) anti-IgA Conjugate****: 1 bottle containing 20 mL of peroxidase labelled rabbit antibody to human IgA; coloured violet, ready to use; black cap.
- **TMB SUBS TMB Substrate Solution**: 1 bottle containing 15 mL 3,3',5,5'-tetra-methyl-benzidine (TMB); ready to use; yellow cap.
- **CONTROL + Varicella-Zoster Virus (VZV) IgA Positive Control*****: 1 bottle containing 2 mL; coloured yellow; ready to use; red cap.
- **CONTROL CO Varicella-Zoster Virus (VZV) IgA Cut-off Control*****: 1 bottle containing 3 mL; coloured yellow; ready to use; green cap.
- **CONTROL - Varicella-Zoster Virus (VZV) IgA Negative Control*****: 1 bottle containing 2 mL; coloured yellow; ready to use; blue cap.

* contains 0.1 % Bronidox L after dilution

** contains 0.2 % Bronidox L

*** contains 0.1 % Kathon

4.2. Materials supplied

- 1 Strip holder
- 1 Cover foil
- 1 Test protocol

4.3. Materials and Equipment needed

- ELISA microwell plate reader, equipped for the measurement of absorbance at 450/620 nm
- Incubator +37°C
- Manual or automatic equipment for rinsing wells
- Pipettes to deliver volumes between 10 and 1000 µL
- Vortex tube mixer
- Deionised or (freshly) distilled water
- Disposable tubes
- Timer

5. STABILITY AND STORAGE

The reagents are stable up to the expiry date stated on the label when stored at 2 - 8 °C.

6. REAGENT PREPARATION

It is very important to bring all reagents, samples and controls to room temperature (20 - 25°C) before starting the test run!

6.1. Coated Snap-off Strips

The ready to use breakapart snap-off strips are coated with Varicella-Zoster Virus (VZV) antigen. Store at 2 - 8°C. *Immediately after removal of strips, the remaining strips should be resealed in the aluminium foil along with the desiccant supplied and stored at 2 - 8 °C; stability until expiry date.*

6.2. Varicella-Zoster Virus (VZV) anti-IgA Conjugate

The bottle contains 20 mL of a solution with anti-human-IgA horseradish peroxidase, buffer, stabilizers, preservatives and an inert violet dye. The solution is ready to use. Store at 2 - 8°C. *After first opening stability until expiry date when stored at 2 - 8°C.*

6.3. Controls

The bottles labelled with Positive, Cut-off and Negative Control contain a ready to use control solution. It contains 0.1% Kathon and has to be stored at 2 - 8°C. *After first opening stability until expiry date when stored at 2 - 8°C.*

6.4. IgA Sample Diluent

The bottle contains 100 mL phosphate buffer, stabilizers, preservatives and an inert yellow dye. It is used for the dilution of the patient specimen. This ready to use solution has to be stored at 2 - 8°C. *After first opening stability until expiry date when stored at 2 - 8°C.*

6.5. Washing Solution (20xconc.)

The bottle contains 50 mL of a concentrated buffer, detergents and preservatives. Dilute washing solution 1+19; e.g. 10 mL washing solution + 190 mL fresh and germ free redistilled water. The diluted buffer is stable for 5 days at room temperature. *Crystals in the solution disappear by warming up to +37 °C in a water bath. After first opening the concentrate is stable until the expiry date.*

6.6. TMB Substrate Solution

The bottle contains 15 mL of a tetramethylbenzidine/hydrogen peroxide system. The reagent is ready to use and has to be stored at 2 - 8°C, away from the light. *The solution should be colourless or could have a slight blue tinge. If the substrate turns into blue, it may have become contaminated and should be thrown away. After first opening stability until expiry date when stored at 2 - 8°C.*

6.7. Stop Solution

The bottle contains 15 mL 0.2 M sulphuric acid solution (R 36/38, S 26). This ready to use solution has to be stored at 2 - 8°C.

After first opening stability until expiry date.

7. SPECIMEN COLLECTION AND PREPARATION

Use human serum or plasma (citrate) samples with this assay. If the assay is performed within 5 days after sample collection, the specimen should be kept at 2 - 8°C; otherwise they should be aliquoted and stored deep-frozen (-20 to -70°C). If samples are stored frozen, mix thawed samples well before testing. *Avoid repeated freezing and thawing.*

Heat inactivation of samples is not recommended.

7.1. Sample Dilution

Before assaying, all samples should be diluted 1+100 with IgA Sample Diluent. Dispense 10µL sample and 1 mL IgA Sample Diluent into tubes to obtain a 1+100 dilution and thoroughly mix with a Vortex.

8. ASSAY PROCEDURE

8.1. Test Preparation

Please read the test protocol carefully **before** performing the assay. Result reliability depends on strict adherence to the test protocol as described. The following test procedure is only validated for manual procedure. If performing the test on ELISA automatic systems we recommend to increase the washing steps from three to five and the volume of washing solution from 300µL to 350µL to avoid washing effects. Prior to commencing the assay, the distribution and identification plan for all specimens and controls should be carefully established on the result sheet supplied in the kit. Select the required number of microtiter strips or wells and insert them into the holder.

Please allocate at least:

- | | |
|----------------------|-----------------------------|
| 1 well (e.g. A1) | for the substrate blank, |
| 1 well (e.g. B1) | for the negative control, |
| 2 wells (e.g. C1+D1) | for the cut-off control and |
| 1 well (e.g. E1) | for the positive control. |

It is recommended to determine controls and patient samples in duplicate, if necessary.

Perform all assay steps in the order given and without any appreciable delays between the steps.

A clean, disposable tip should be used for dispensing each control and sample.

Adjust the incubator to $+37^{\circ} \pm 1^{\circ}\text{C}$.

1. Dispense 100 μL controls and diluted samples into their respective wells. Leave well A1 for substrate blank.
2. Cover wells with the foil supplied in the kit.
3. **Incubate for 1 hour ± 5 min at $+37 \pm 1^{\circ}\text{C}$.**
4. When incubation has been completed, remove the foil, aspirate the content of the wells and wash each well three times with 300 μL of Washing Solution. Avoid overflows from the reaction wells. The soak time between each wash cycle should be >5 sec. At the end carefully remove remaining fluid by tapping strips on tissue paper prior to the next step!

Note: Washing is critical! Insufficient washing results in poor precision and falsely elevated absorbance values.

5. Dispense 100 μL Varicella-Zoster Virus (VZV) anti-IgA Conjugate into all wells except for the blank well (e.g. A1). Cover with foil.
6. **Incubate for 30 min at room temperature. Do not expose to direct sunlight.**
7. Repeat step 4.
8. Dispense 100 μL TMB Substrate Solution into all wells
9. **Incubate for exactly 15 min at room temperature in the dark.**
10. Dispense 100 μL Stop Solution into all wells in the same order and at the same rate as for the TMB Substrate Solution.

Any blue colour developed during the incubation turns into yellow.

Note: Highly positive patient samples can cause dark precipitates of the chromogen! These precipitates have an influence when reading the optical density. Predilution of the sample with physiological sodium chloride solution, for example 1+1, is recommended. Then dilute the sample 1+100 with dilution buffer and multiply the results in U by 2.

11. Measure the absorbance of the specimen at 450/620 nm within 30 min after addition of the Stop Solution.

8.2. Measurement

Adjust the ELISA Microwell Plate Reader **to zero** using the **substrate blank in well A1**.

If - due to technical reasons - the ELISA reader cannot be adjusted to zero using the substrate blank in well A1, subtract the absorbance value of well A1 from all other absorbance values measured in order to obtain reliable results!

Measure the absorbance of all wells at **450 nm** and record the absorbance values for each control and patient sample in the distribution and identification plan.

Dual wavelength reading using 620 nm as reference wavelength is recommended.

Where applicable calculate the **mean absorbance values** of all duplicates.

9. RESULTS

9.1. Run Validation Criteria

In order for an assay to be considered valid, the following criteria must be met:

- **Substrate blank** in A1: Absorbance value **< 0.100**.
- **Negative control** in B1: Absorbance value **< 0.200 and < cut-off**
- **Cut-off control** in C1 and D1: Absorbance value **0.150 – 1.30**.
- **Positive control** in E1: Absorbance value **> cut-off**.

If these criteria are not met, the test is not valid and must be repeated.

9.2. Calculation of Results

The cut-off is the mean absorbance value of the Cut-off control determinations.

Example: Absorbance value Cut-off control 0.44 + absorbance value Cut-off control 0.42 = 0.86 / 2 = 0.43

Cut-off = 0.43

9.3. Interpretation of Results

Samples are considered **POSITIVE** if the absorbance value is higher than 10% over the cut-off.

Samples with an absorbance value of 10% above or below the cut-off should not be considered as clearly positive or negative

→ **grey zone**

It is recommended to repeat the test again 2 - 4 weeks later with a fresh sample. If results in the second test are again in the grey zone the sample has to be considered **NEGATIVE**.

Samples are considered **NEGATIVE** if the absorbance value is lower than 10% below the cut-off.

9.3.1. Results in Units

$$\frac{\text{Patient (mean) absorbance value} \times 10}{\text{Cut-off}} = [\text{Units} = \text{U}]$$

$$\text{Example: } \frac{1.786 \times 10}{0.43} = 42 \text{ U (Units)}$$

Cut-off:	10	U
Grey zone:	9-11	U
Negative:	<9	U
Positive:	>11	U

10. SPECIFIC PERFORMANCE CHARACTERISTICS

10.1. Precision

Interassay	n	Mean	Cv (%)
Pos. Serum	7	0.65	5.3

Intraassay	n	Mean	Cv (%)
Pos. Serum	3	0.68	3.8

10.2. Diagnostic Specificity

The diagnostic specificity is defined as the probability of the assay of scoring negative in the absence of the specific analyte.

It is > 90 %.

10.3. Diagnostic Sensitivity

The diagnostic sensitivity is defined as the probability of the assay of scoring positive in the presence of the specific analyte.

It is > 90 %.

10.4. Interferences

Interferences with hemolytic, lipemic or icteric sera are not observed up to a concentration of 10 mg/mL hemoglobin, 5 mg/mL triglycerides and 0.2 mg/mL bilirubin.

Note: The results refer to the groups of samples investigated; these are not guaranteed specifications.

11. LIMITATIONS OF THE PROCEDURE

Bacterial contamination or repeated freeze-thaw cycles of the specimen may affect the absorbance values. Diagnosis of an infectious disease should not be established on the basis of a single test result. A precise diagnosis should take into consideration clinical history, symptomatology as well as serological data.

In immunocompromised patients and newborns serological data only have restricted value.

12. PRECAUTIONS AND WARNINGS

- In compliance with article 1 paragraph 2b European directive 98/79/EC the use of the in vitro diagnostic medical devices is intended by the manufacturer to secure suitability, performances and safety of the product. Therefore the test procedure, the information, the precautions and warnings in the instructions for use have to be strictly followed. The use of the testkits with analyzers and similar equipment has to be validated. Any change in design, composition and test procedure as well as for any use in combination with other products not approved by the manufacturer is not authorized; the user himself is responsible for such changes. The manufacturer is not liable for false results and incidents for these reasons. The manufacturer is not liable for any results by visual analysis of the patient samples.
- Only for in-vitro diagnostic use.
- All components of human origin used for the production of these reagents have been tested for anti-HIV antibodies, anti-HCV antibodies and HBsAg and have been found to be non-reactive. Nevertheless, all materials should still be regarded and handled as potentially infectious.
- Do not interchange reagents or strips of different production lots.
- No reagents of other manufacturers should be used along with reagents of this test kit.
- Do not use reagents after expiry date stated on the label.
- Use only clean pipette tips, dispensers, and lab ware.
- Do not interchange screw caps of reagent vials to avoid cross-contamination.
- Close reagent vials tightly immediately after use to avoid evaporation and microbial contamination.
- After first opening and subsequent storage check conjugate and control vials for microbial contamination prior to further use.
- To avoid cross-contamination and falsely elevated results pipette patient samples and dispense conjugate without splashing accurately to the bottom of wells.
- The ELISA is only designed for qualified personnel who are familiar with good laboratory practice.

WARNING: In the used concentration Bronidox L has hardly any toxicological risk upon contact with skin and mucous membranes!

WARNING: Sulphuric acid irritates eyes and skin. Keep out of the reach of children. Upon contact with the eyes, rinse thoroughly with water and consult a doctor!

12.1. Disposal Considerations

Residues of chemicals and preparations are generally considered as hazardous waste. The disposal of this kind of waste is regulated through national and regional laws and regulations. Contact your local authorities or waste management companies which will give advice on how to dispose hazardous waste.

13. ORDERING INFORMATION

Prod. No.: 30114080 Varicella zoster virus IgA ELISA (96 Determinations)

BIBLIOGRAPHY / LITERATUR / BIBLIOGRAPHIE / BIBLIOGRAFIA / BIBLIOGRAFÍA

- Croen, K.D., S.E. Straus: VZV latency. Annu. Rev. Microbiol. 45 (1991) 265-285
- Slugosch, D., A.M. Eis-Hübingen, J-P. Klein et al.: Diagnosis of acute and latent VZV infections using polymerase chain reaction. J. Med. Virol. 35 (1991) 136-141
- Enders, G., VZV infections in pregnancy. Prog. med. Virol. 25 (1984) 166-196
- Harper, D.R., H.O. Kangro, R.B. Health: Serological responses in varicella and zoster assayed by immunoblotting. J. Med. Virol. 25 (1988) 387-398
- Health, R.B., H.O. Kangro Varizella zoster. Inc: Zuckermann, A.J., J.E. Banatvala J.R. Pattison (eds.): Principles and Practice of Clinical Virology, 2nd edition. John Wiley and Sons, Chichester-New

SCHEME OF THE ASSAY
Varicella zoster virus IgA ELISA

Test Preparation

Prepare reagents and samples as described.
 Establish the distribution and identification plan for all specimens and controls on the result sheet.
 Select the required number of microtiter strips or wells and insert them into the holder.

Assay Procedure

	Substrate blank (e.g. A1)	Negative control	Positive control	Cut-off control	Sample (diluted 1+100)
Negative control	-	100µL	-	-	-
Positive control	-	-	100µL	-	-
Cut-off control	-	-	-	100µL	-
Sample (diluted 1+100)	-	-	-	-	100µL
Cover wells with foil supplied in the kit Incubate for 1 h at +37°C					
Wash each well three times with 300µL of washing solution					
Conjugate	-	100µL	100µL	100µL	100µL
Cover wells with foil supplied in the kit Incubate for 30 min at room temperature					
Wash each well three times with 300µL of washing solution					
TMB Substrate	100µL	100µL	100µL	100µL	100µL
Incubate for exactly 15 min at room temperature in the dark					
Stop Solution	100µL	100µL	100µL	100µL	100µL
Photometric measurement at 450 nm (reference wavelength: 620 nm)					

1. EINLEITUNG

Das Varizella-Zoster Virus (VZV) aus der Familie der Herpesviridae ist neben dem Herpes-simplex-Virus 1 und 2 das dritte humanpathogene Alpha-Herpesvirus. Es besteht aus einem Nukleokapsid, das einen Kern mit einer linearen Doppelstrang-DNS umschließt. Ein proteinhaltiges Tegument trennt das Kapsid von der Fetthülle, das die wichtigsten viralen Glykoproteine enthält. Das Virus kann zwei verschiedene klinische Krankheitsbilder verursachen: Varizellen (Windpocken) bei exogener Neuinfektion und Herpes zoster (Gürtelrose) bei endogener Reaktivierung.

Die Eintrittspforten in den menschlichen Körper sind die Schleimhaut des oberen Respirationstraktes und die Konjunktiven. Über die regionalen Lymphknoten erreicht das Virus Milz und Leber. Von hier aus breitet sich das Virus über infizierte mononukleäre Zellen mukokutan aus. Die Infektion epidermaler Zellen endet durch ausgeprägte zytopathogene Effekte in den typischen makulopapulären Hautläsionen. In dieser Phase werden auch die Zellen der Lumbosakralganglien infiziert. Viele Jahre später (typischerweise nach dem 45. Lebensjahr) kann es zur Reaktivierung des Virus mit Entzündung des befallenen Ganglions kommen. Typisch ist dabei die scharf begrenzte, einseitige Lokalisation der sehr schmerhaften Läsionen der Haut im Bereich der sie versorgenden sensiblen Nerven.

Varizellen sind äußerst kontagiös; nach einer Exposition würden über 90 von 100 empfänglichen, d. h. seronegativen Personen erkranken (Kontagionsindex nahe 1.0). Der Durchseuchungsgrad beträgt im Erwachsenenalter etwa 80 - 90 %. Das Virus kommt endemisch in der Bevölkerung vor und wird vor allem auch im Zuge saisonaler Häufungen - in gemäßigten Breitengraden im Winter und Frühjahr - übertragen. Die Übertragung erfolgt aerogen durch virushaltige Tröpfchen, die beim Atmen oder Husten ausgeschieden werden (und u. U. im Umkreis von mehreren Metern zur Ansteckung führen können). Ferner ist eine Übertragung durch virushaltigen Bläscheninhalt oder Krusten als Schmierinfektion möglich. Bei Herpes zoster besteht eine geringere Kontagiosität. Eine diaplazentare Übertragung ist selten, kann aber in etwa 1 % der Varizellenerkrankungen bei Schwangeren zum kongenitalen Varizellensyndrom führen, sofern die Erkrankung vor der 21. Schwangerschaftswoche aufgetreten ist. Neugeborene, immuninkompetente Personen und Patienten unter einer Glukokortikoidtherapie sind in der Regel durch schwere Krankheitsverläufe besonders gefährdet. Es existiert die Möglichkeit der aktiven und passiven Immunisierung.

Spezies	Erkrankung	Symptome	Infektionsmodus
Varizella-Zoster Virus (VZV)	Windpocken	Fieber, Exanthem mit Juckreiz und Bläschen	aerogen durch virushaltige Tröpfchen, die beim Atmen und Husten ausgeschieden werden
	Herpes zoster (Gürtelrose)	Komplikationen: Pneumonie, Meningitis, Enzephalitis	Schmierinfektion

Nachweis

- Mikroskopie
- Serologie: Nachweis spezifischer Antikörper mittels ELISA

2. VERWENDUNGSZWECK

Der Varicella zoster virus IgA ELISA ist für den qualitativen Nachweis spezifischer IgA-Antikörper gegen Varizellen in humanem Serum oder Plasma (Citrat) bestimmt.

3. TESTPRINZIP

Die qualitative immunenzymatische Bestimmung von spezifischen IgA-Antikörpern gegen das Varizella-Zoster Virus beruht auf der ELISA (Enzyme-linked Immunosorbent Assay)-Technik.

Mikrotiterstreifen als solide Phase sind beschichtet mit Varizella-Zoster Virus spezifischen Antigenen. Vorhandene spezifische Antikörper in der Probe binden an die immobilisierten Antigene der Mikrotiterplatte. Meerrettich-Peroxidase (HRP)-konjugierte anti-human-IgA Antikörper binden an Antigen-Antikörperkomplexe in positiven Proben. Die entstandenen Immunkomplexe werden durch Blaufärbung nach Inkubation mit Tetramethylbenzidin (TMB) -Substratlösung nachgewiesen. Stoppen der enzymatischen Reaktion mit Schwefelsäure führt zu einem Farbumschlag von blau zu gelb, der einfach nachgewiesen und mit einem ELISA-Reader bei 450 nm gemessen werden kann.

4. MATERIALIEN

4.1. Mitgelieferte Reagenzien

- **MTP Varizella-Zoster Virus beschichtete Mikrotiterstreifen (IgA):** 12 teilbare 8er-Streifen, beschichtet mit Varizella-Zoster Virus-Antigen; in wiederverschließbarem Aluminiumbeutel.
- **SAMPLEDIL IgA-Probenverdünnungspuffer***:** 1 Flasche mit 100 mL Puffer zur Probenverdünnung; pH 7.2 ± 0.2; gelb gefärbt; gebrauchsfertig; weiße Verschlusskappe.
- **TMB STOP Stopplösung:** 1 Fläschchen mit 15 mL Schwefelsäure, 0.2 mol/L, gebrauchsfertig; rote Verschlusskappe.
- **WASHBUF CONC Waschlösung (20x konz.)*:** 1 Flasche mit 50 mL eines 20-fach konzentrierten Puffers zum Waschen der Kavitäten; pH 7.2 ± 0.2; weiße Verschlusskappe.
- **ENZCONJ Varizella-Zoster Virus anti-IgA-Konjugat**:** 1 Flasche mit 20 mL Peroxidase-konjugierten Antikörpern gegen humanes IgA; violett gefärbt; gebrauchsfertig; schwarze Verschlusskappe.
- **TMB SUBS TMB-Substratlösung:** 1 Fläschchen mit 15 mL 3,3',5,5'-Tetramethylbenzidin (TMB); gebrauchsfertig; gelbe Verschlusskappe.
- **CONTROL + Varizella-Zoster Virus IgA Positivkontrolle***:** 1 Fläschchen mit 2 mL; gelb gefärbt; rote Verschlusskappe; gebrauchsfertig.
- **CONTROL CO Varizella-Zoster Virus IgA Cut-off Kontrolle***:** 1 Fläschchen mit 3 mL; gelb gefärbt; grüne Verschlusskappe; gebrauchsfertig.
- **CONTROL - Varizella-Zoster Virus IgA Negativkontrolle***:** 1 Fläschchen mit 2 mL; gelb gefärbt; blaue Verschlusskappe; gebrauchsfertig.

* enthält 0.1 % Bronidox L nach Verdünnung

** enthält 0.2 % Bronidox L

*** enthält 0.1 % Kathon

4.2. Mitgeliefertes Zubehör

- 1 selbstklebende Abdeckfolie
- 1 Rahmenhalter
- 1 Arbeitsanleitung

4.3. Erforderliche Materialien und Geräte

- Photometer mit Filtern 450/620 nm
- Feuchtkammer/Brutschrank mit Thermostat
- Manuelle oder automatische Waschvorrichtung
- Mikropipetten mit Einmalspitzen (10, 100, 200, 1000 µL)
- Vortex-Mischer
- Plastikröhrchen für den einmaligen Gebrauch
- Röhrchen-Ständer
- Aqua dest.
- Timer

5. STABILITÄT UND LAGERUNG

Testkit bei 2 - 8°C lagern. Die Reagenzien nicht nach den angegebenen Verfallsdaten verwenden. Die Verfallsdaten sind jeweils auf den Flaschenetiketten und auf dem Außenetikett angegeben.

6. VORBEREITUNG DER REAGENZIEN

Alle Reagenzien, Proben und Kontrollen sind vor ihrer Verwendung auf Raumtemperatur (20 - 25°C) zu bringen!

6.1. Beschichtete Streifen

Die abbrechbaren Streifen sind mit inaktiviertem Varizella-Zoster Virus Antigen beschichtet. Die gebrauchsfertigen Vertiefungen sind bei 2 - 8°C aufzubewahren. Nichtverbrauchte Vertiefungen im Aluminiumbeutel zusammen mit dem Trockenmittel sofort wieder verschließen und bei 2 - 8°C lagern. Haltbarkeit bis zum angegebenen Verfallsdatum.

6.2. Varizella-Zoster Virus anti-IgA-Konjugat

Die Flasche enthält 20 mL einer Lösung von anti-human IgA-Meerrettichperoxidase, Puffer, Stabilisatoren, Konservierungsmittel und einen inerten violetten Farbstoff. Die gebrauchsfertige Lösung ist bei 2 - 8°C aufzubewahren. *Nach dem ersten Öffnen haltbar bis zum angegebenen Verfallsdatum (bei 2 - 8°C).*

6.3. Kontrollen

Die Fläschchen mit Kontrollen enthalten 2 bzw. 3 mL gebrauchsfertige Kontrolllösung. Die gebrauchsfertigen Lösungen sind bei 2 - 8°C aufzubewahren und enthalten 0.1 % Kathon. *Nach dem ersten Öffnen haltbar bis zum angegebenen Verfallsdatum (bei 2 - 8°C).*

6.4. IgA-Probenverdünnungspuffer

Die Flasche enthält 100 mL Phosphatpuffer, Stabilisatoren, Konservierungsmittel und einen inerten gelben Farbstoff. Die gebrauchsfertige Lösung ist bei 2 - 8°C aufzubewahren. Die Lösung wird für die Verdünnung der Proben eingesetzt. *Nach dem ersten Öffnen haltbar bis zum angegebenen Verfallsdatum (bei 2 - 8°C).*

6.5. Waschlösung (20x konz.)

Die Flasche enthält 50 mL konzentrierten Puffer, Detergenzien und Konservierungsmittel. Der Inhalt wird auf einen Liter mit Aqua dest. verdünnt (1+19). Der verdünnte Puffer ist bei Raumtemperatur 5 Tage haltbar. Die Waschlösung wird zum Waschen der Streifen eingesetzt. *Sollte eine Kristallisation im Konzentrat auftreten, die Waschlösung auf +37°C erwärmen und vor dem Verdünnen gut mischen. Nach dem ersten Öffnen, Konzentrat haltbar bis zum angegebenen Verfallsdatum (bei 2 - 8°C).*

6.6. TMB-Substratlösung

Das Fläschchen enthält 15 mL eines Tetramethylbenzidin/Wasserstoffperoxidgemisches. Die gebrauchsfertige Lösung ist bei 2 - 8°C vor Licht geschützt aufzubewahren *Die Lösung ist leicht hellblau. Sollte die TMB-Substratlösung dunkelblau sein, ist sie kontaminiert und kann nicht im Test verwendet werden. Nach dem ersten Öffnen haltbar bis zum Verfallsdatum bei sachgerechter Lagerung von 2 - 8°C.*

6.7. Stopplösung

Das Fläschchen enthält 15 mL 0.2 M Schwefelsäure (R36/38, S26). Die gebrauchsfertige Lösung ist bei 2 - 8°C aufzubewahren. *Nach dem ersten Öffnen haltbar bis zum angegebenen Verfallsdatum (bei 2 - 8°C).*

7. ENTNAHME UND VORBEREITUNG DER PROBEN

Es sollten humane Serum- oder Plasmaproben (Citrat) verwendet werden. Werden die Bestimmungen innerhalb von 5 Tagen nach Blutentnahme durchgeführt, können die Proben bei 2 - 8°C aufbewahrt werden, sonst tiefgefrieren (-70 bis -20°C). Wiederaufgetaute Proben vor dem Verdünnen gut schütteln. *Wiederholtes Tiefgefrieren und Auftauen vermeiden!* Hitzeinaktivierung der Proben wird nicht empfohlen.

7.1. Probenverdünnung

Proben vor Testbeginn im Verhältnis 1+100 mit IgA-Probenverdünnungspuffer verdünnen, z.B. 10µL Probe und 1 mL IgA-Probenverdünnungspuffer in die entsprechenden Röhrchen pipettieren, um eine Verdünnung von 1+100 zu erhalten; gut mischen (Vortex).

8. TESTDURCHFÜHRUNG

8.1. Testvorbereitung

Gebrauchsinformation **vor** Durchführung des Tests sorgfältig lesen. Für die Zuverlässigkeit der Ergebnisse ist es notwendig, die Arbeitsanleitung genau zu befolgen. Die folgende Testdurchführung ist für die manuelle Methode validiert. Beim Arbeiten mit ELISA Automaten empfehlen wir, um Wascheffekte auszuschließen, die Zahl der Waschschrifte von drei auf fünf und das Volumen der Waschlösung von 300 µL auf 350 µL zu erhöhen. Vor Testbeginn auf dem mitgelieferten Ergebnisblatt die Verteilung bzw. Position der Patientenproben und Standards auf den Mikrotiterstreifen genau festlegen. Die benötigte Anzahl von Mikrotiterstreifen (Kavitäten) in den Streifenhalter einsetzen.

Hierbei mindestens

1 Vertiefung	(z.B. A1)	für den Substratleerwert (Blank),
1 Vertiefungen	(z.B. B1)	für die Negativ Kontrolle und
2 Vertiefungen	(z.B. C1+D1)	für die Cut-off Kontrolle und
1 Vertiefung	(z.B. E1)	für die Positiv Kontrolle vorsehen

Prinzipien der Qualitätssicherung in der Laboratoriumsmedizin erfordern zur höheren Sicherheit für Kontrollen und Patientenproben mindestens Doppelbestimmungen.

Den Test in der angegebenen Reihenfolge und ohne Verzögerung durchführen.

Für jeden Pipettierschritt der Kontrollen und Proben saubere Einmalspitzen verwenden.

Den Brutschrank auf $+37 \pm 1^\circ\text{C}$ einstellen.

1. Je 100 μL Kontrollen und vorverdünnte Proben in die entsprechenden Vertiefungen pipettieren. Vertiefung A1 ist für den Substratleerwert vorgesehen.
2. Die Streifen mit der mitgelieferten Abdeckfolie bedecken.
3. **1 h ± 5 min bei $+37^\circ\text{C}$ inkubieren.**
4. Am Ende der Inkubationszeit Abdeckfolie entfernen und die Inkubationsflüssigkeit aus den Teststreifen absaugen. Anschließend dreimal mit 300 μL Waschlösung waschen. Überfließen von Flüssigkeit aus den Vertiefungen vermeiden. Intervall zwischen Waschen und Absaugen sollte mindestens 5 sec betragen. Nach dem Waschen die Teststreifen mit den Öffnungen nach unten kurz auf Fliesspapier ausklopfen, um die restliche Flüssigkeit zu entfernen.

Beachte: *Der Waschvorgang ist wichtig, da unzureichendes Waschen zu schlechter Präzision und falsch erhöhten Messergebnissen führt!*

5. 100 μL Varizella-Zoster Virus anti-IgA-Konjugat in alle Vertiefungen, mit Ausnahme der für die Berechnung des Leerwertes vorgesehenen, pipettieren. Mit Folie abdecken.
6. **30 min bei Raumtemperatur ($20 - 25^\circ\text{C}$) inkubieren. Nicht dem direkten Sonnenlicht aussetzen.**
7. Waschvorgang gemäß Punkt 4 wiederholen.
8. 100 μL TMB-Substratlösung in alle Vertiefungen pipettieren.
9. **Genau 15 min im Dunkeln bei Raumtemperatur ($20 - 25^\circ\text{C}$) inkubieren.**
10. In alle Vertiefungen 100 μL Stopplösung in der gleichen Reihenfolge und mit den gleichen Zeitintervallen wie bei der TMB-Substratlösung Zugabe pipettieren. *Während der Inkubation gebildete blaue Farbe schlägt in gelb um.*

Hinweis: *Hochpositive Patientenproben können schwärzliche Präzipitate des Chromogens verursachen! Diese Präzipitate beeinflussen die Messwerte. Es wird empfohlen, die Patientenprobe mit physiologischer Kochsalzlösung 1 + 1 zu verdünnen und anschließend die verdünnte Probe mit IgA-Probenverdünnungspuffer 1 + 100 für den Test vorzubereiten. Das Ergebnis in U wird in diesem Fall mit zwei multipliziert.*

11. Die Extinktion der Lösung in jeder Vertiefung bei 450/620 nm innerhalb von 30 min nach Zugabe der Stopplösung messen

8.2. Messung

Mit Hilfe des Substratleerwertes (Blank) **in A1 den Nullabgleich** des Mikrotiterplatten-Photometers (ELISA-Readers) vornehmen.

Falls diese Eichung aus technischen Gründen nicht möglich ist, muss nach der Messung der Extinktionswert der Position A1 von allen anderen Extinktionswerten abgezogen werden, um einwandfreie Ergebnisse zu erzielen!

Extinktion aller Kavitäten bei **450 nm** messen und die Messwerte der Kontrollen und Proben in das Ergebnisblatt eintragen.

Eine bichromatische Messung mit der Referenzwellenlänge 620 nm wird empfohlen.

Falls Doppel- oder Mehrfachbestimmungen durchgeführt wurden, den **Mittelwert der Extinktionswerte** berechnen.

9. BERECHNUNG DER ERGEBNISSE

9.1. Testgültigkeitskriterien

Der Test wurde richtig durchgeführt, wenn er folgende Kriterien erfüllt:

- **Substrat-Leerwert** in A1: **Extinktion < 0.100**

- **Negativ Kontrolle** in B1: Extinktion < **0.200 und < cut-off**
- **Cut-off Kontrolle** in C1 und D1: Extinktionswerte **0.150 – 1.300**
- **Positiv Kontrolle** in E1: Extinktionswerte > **Cut-off**

Sind diese Kriterien nicht erfüllt, ist der Testlauf ungültig und muss wiederholt werden.

9.2. Messwertberechnung

Der Cut-off ergibt sich aus dem Mittelwert der gemessenen Extinktionen der beiden Cut-off Kontrollen.

Beispiel: $0.44 \text{ OD Cut-off Kontrolle} + 0.42 \text{ OD Cut-off Kontrolle} = 0.86 / 2 = \underline{\underline{0.43}}$

Cut-off = 0.43

9.3. Interpretation der Ergebnisse

Patientenproben gelten als **positiv**, wenn der Extinktionswert mindestens 10 % höher liegt als der Cut-Off.

Patientenproben mit Extinktionswerten 10 % über bzw. unter dem Cut-Off können nicht eindeutig als positiv bzw. negativ angesehen werden → **Grauzone**

Es wird empfohlen den Test nach 2 bis 4 Wochen mit einer frischen Patientenprobe zu wiederholen. Finden sich die Ergebnisse erneut innerhalb der Grauzone, gilt die Probe als **negativ**.

Patientenproben gelten als **negativ**, wenn der Extinktionswert mindestens 10 % unterhalb des Cut-Offs liegt.

9.3.1. Ergebnisse in Einheiten [U]

$$\frac{\text{Mittlere Extinktion der Patientenprobe} \times 10}{\text{Cut-Off}} = [\text{Einheiten} = \text{units} = U]$$

$$\text{Beispiel: } \frac{1.786 \times 10}{0.43} = 42 \text{ U (Units)}$$

Cut-Off:	10	U
Grauzone:	9-11	U
Negativ:	<9	U
Positiv:	>11	U

10. TESTMERKMALE

10.1. Präzision

Interassay	n	Mittelwert	Vk (%)
Pos. Serum	7	0.65	5.3

Intraassay	n	Mittelwert	Vk (%)
Pos. Serum	3	0.68	3.8

10.2. Diagnostische Spezifität

Die diagnostische Spezifität ist definiert als die Wahrscheinlichkeit des Tests, ein negatives Ergebnis bei Fehlen des spezifischen Analyten zu liefern. Sie beträgt >90%.

10.3. Diagnostische Sensitivität

Die diagnostische Sensitivität ist definiert als die Wahrscheinlichkeit des Tests, ein positives Ergebnis bei Vorhandensein des spezifischen Analyten zu liefern. Sie beträgt >90%.

10.4. Interferenzen

Hämolytische, lipämische und ikterische Proben ergaben bis zu einer Konzentration von 10 mg/mL für Hämoglobin, von 5 mg/mL Triglyceride und von 0.2 mg/mL für Bilirubin keine Interferenzen im vorliegenden ELISA.

Hinweis: Die Ergebnisse beziehen sich auf die untersuchten Probenkollektive; es handelt sich nicht um garantierte Spezifikationen.

11. GRENZEN DES VERFAHRENS

Kontamination der Proben durch Bakterien oder wiederholtes Einfrieren und Auftauen können zu einer Veränderung der Messwerte führen. Die Diagnose einer Infektionskrankheit darf nicht allein auf der Basis des Ergebnisses einer Bestimmung gestellt werden. Die anamnestischen Daten sowie die Symptomatologie des Patienten müssen zusätzlich zu den serologischen Ergebnissen in Betracht gezogen werden. Bei Immunsupprimierten und Neugeborenen besitzen die Ergebnisse der serologischen Tests nur einen begrenzten Wert.

12. SICHERHEITSMASSNAHMEN UND WARNHINWEISE

- Gemäß Art. 1 Abs. 2b der EU-Richtlinie 98/79/EG legt der Hersteller die Zweckbestimmung von In-vitro-Diagnostika fest, um deren Eignung, Leistung und Sicherheit sicherzustellen. Daher sind die Testdurchführung, die Information, die Sicherheitsmaßnahmen und Warnhinweise in der Gebrauchsanweisung strikt zu befolgen. Bei Anwendung des Testkits auf Diagnostika-Geräten ist die Testmethode zu validieren. Jede Änderung am Aussehen, der Zusammensetzung und der Testdurchführung sowie jede Verwendung in Kombination mit anderen Produkten, die der Hersteller nicht autorisiert hat, ist nicht zulässig; der Anwender ist für solche Änderungen selbst verantwortlich. Der Hersteller haftet für falsche Ergebnisse und Vorkommnisse aus solchen Gründen nicht. Auch für falsche Ergebnisse aufgrund von visueller Auswertung wird keine Haftung übernommen.
- Nur für in-vitro-Diagnostik.
- Alle verwendeten Bestandteile menschlichen Ursprungs sind auf Anti-HIV-AK, Anti-HCV-AK und HBsAG nicht-reakтив getestet. Dennoch sind alle Materialien als potentiell infektiös anzusehen und entsprechend zu behandeln.
- Reagenzien und Mikrotiterplatten unterschiedlicher Chargen nicht untereinander austauschen.
- Keine Reagenzien anderer Hersteller zusammen mit den Reagenzien dieses Testkits verwenden.
- Nicht nach Ablauf des Verfallsdatums verwenden.
- Nur saubere Pipettenspitzen, Dispenser und Labormaterialien verwenden.
- Verschlusskappen der einzelnen Reagenzien nicht untereinander vertauschen.
- Flaschen sofort nach Gebrauch fest verschließen, um Verdunstung und mikrobielle Kontamination zu vermeiden.
- Nach dem ersten Öffnen Konjugat- und Standardfläschchen vor weiterem Gebrauch auf mikrobielle Kontamination prüfen.
- Zur Vermeidung von Kreuzkontamination und falsch erhöhten Resultaten Patientenproben und Konjugat sorgfältig in die Kavitäten pipettieren.
- Der ELISA ist nur für die Anwendung durch Fachpersonal vorgesehen, welches die Arbeitstechniken einwandfrei beherrscht.

WARNUNG: Bronidox L zeigt in der verwendeten Konzentration nahezu keine toxikologischen Risiken an Haut bzw. Schleimhaut.

WARNUNG: Schwefelsäure reizt Augen und Haut! Nach Berührung mit den Augen gründlich mit Wasser spülen und einen Arzt aufsuchen.

12.1. Entsorgungshinweise

Chemikalien und Zubereitungen sind in der Regel Sonderabfälle. Deren Beseitigung unterliegt den nationalen abfallrechtlichen Gesetzen und Verordnungen. Die zuständige Behörde informiert über die Entsorgung von Sonderabfällen.

13. BESTELLINFORMATIONEN

Produktnummer: 30114080 Varicella zoster virus IgA ELISA (96 Bestimmungen)

BIBLIOGRAPHY / LITERATUR / BIBLIOGRAPHIE / BIBLIOGRAFIA / BIBLIOGRAFÍA

- Croen, K.D., S.E. Straus: VZV latency. Annu. Rev. Microbiol. 45 (1991) 265-285
 Drugosch, D., A.M. Eis-Hübingen, J-P. Klein et al.: Diagnosis of acute and latent VZV infections using polymerase chain reaction. J. Med. Virol. 35 (1991) 136-141
 Enders, G., VZV infections in pregnancy. Prog. med. Virol. 25 (1984) 166-196
 Harper, D.R., H.O. Kangro, R.B. Health: Serological responses in varicella and zoster assayed by immunoblotting. J. Med. Virol. 25 (1988) 387-398
 Health, R.B., H.O. Kangro Varizella zoster. Inc: Zuckermann, A.J., J.E. Banatvala J.R. Pattison (eds.): Principles and Practice of Clinical Virology, 2nd edition. John Wiley and Sons, Chichester-New

SCHEME OF THE ASSAY
 Varicella zoster virus IgA ELISA

Test Preparation

Prepare reagents and samples as described.
 Establish the distribution and identification plan for all specimens and controls on the result sheet.
 Select the required number of microtiter strips or wells and insert them into the holder.

Assay Procedure

	Substrate blank (e.g. A1)	Negative control	Positive control	Cut-off control	Sample (diluted 1+100)
Negative control	-	100µL	-	-	-
Positive control	-	-	100µL	-	-
Cut-off control	-	-	-	100µL	-
Sample (diluted 1+100)	-	-	-	-	100µL
Cover wells with foil supplied in the kit Incubate for 1 h at +37°C					
Wash each well three times with 300µL of washing solution					
Conjugate	-	100µL	100µL	100µL	100µL
Cover wells with foil supplied in the kit Incubate for 30 min at room temperature					
Wash each well three times with 300µL of washing solution					
TMB Substrate	100µL	100µL	100µL	100µL	100µL
Incubate for exactly 15 min at room temperature in the dark					
Stop Solution	100µL	100µL	100µL	100µL	100µL
Photometric measurement at 450 nm (reference wavelength: 620 nm)					

Symbols / Symbole / Symbôles / Símbolos / Símbolos / Σύμβολα

	Cat.-No.: / Kat.-Nr.: / No.- Cat.: / Cat.-No.: / N.º Cat.: / N.-Cat.: / Αριθμός-Κατ.:
	Lot-No.: / Chargen-Bez.: / No. Lot: / Lot-No.: / Lote N.º: / Lotto n.: / Αριθμός -Παραγωγή:
	Use by: / Verwendbar bis: / Utiliser à: / Usado por: / Usar até: / Da utilizzare entro: / Χρησιμοποιείται από:
	No. of Tests: / Kitgröße: / Nb. de Tests: / No. de Determ.: / N.º de Testes: / Quantità dei tests: / Αριθμός εξετάσεων:
	Concentrate / Konzentrat / Concentré / Concentrar / Concentrado / Concentrato / Συμπύκνωμα
	Lyophilized / Lyophilisat / Lyophilisé / Liofilizado / Liofilizzato / Λυοφιλισμένο
	In Vitro Diagnostic Medical Device. / In-vitro-Diagnostikum. / Appareil Médical pour Diagnostics In Vitro. / Dispositivo Médico para Diagnóstico In Vitro. / Equipamento Médico de Diagnóstico In Vitro. / Dispositivo Medico Diagnóstico In vitro. / Ιατρική συσκευή για In-Vitro Διάγνωση.
	Evaluation kit. / Nur für Leistungsbewertungszwecke. / Kit pour évaluation. / Juego de Reactivos para Evaluació. / Kit de avaliação. / Kit di evaluazione. / Κιτ Αξιολόγησης.
	Read instructions before use. / Arbeitsanleitung lesen. / Lire la fiche technique avant emploi. / Lea las instrucciones antes de usar. / Ler as instruções antes de usar. / Leggere le istruzioni prima dell'uso. / Διαβάστε τις οδηγίες πριν την χρήση.
	Keep away from heat or direct sun light. / Vor Hitze und direkter Sonneneinstrahlung schützen. / Garder à l'abri de la chaleur et de toute exposition lumineuse. / Manténgase alejado del calor o la luz solar directa. / Manter longe do calor ou luz solar directa. / Non esporre ai raggi solari. / Να φυλάσσεται μακριά από θερμότητα και άμεση επαφή με το φως του ηλίου.
	Store at: / Lagern bei: / Stocker à: / Almacene a: / Armazenar a: / Conservare a: / Αποθήκευση στους:
	Manufacturer: / Hersteller: / Fabricant: / Productor: / Fabricante: / Fabbricante: / Παραγωγός:
	Caution! / Vorsicht! / Attention! / ¡Precaución! / Cuidado! / Attenzione! / Προσοχή!

Symbols of the kit components see MATERIALS SUPPLIED.

Die Symbole der Komponenten sind im Kapitel KOMPONENTEN DES KITS beschrieben.

Voir MATERIEL FOURNI pour les symboles des composants du kit.

Símbolos de los componentes del juego de reactivos, vea MATERIALES SUMINISTRADOS.

Para símbolos dos componentes do kit ver MATERIAIS FORNECIDOS.

Per i simboli dei componenti del kit si veda COMPONENTI DEL KIT.

Για τα σύμβολα των συστατικών του κιτ συμβουλευτείτε το ΠΑΡΕΧΟΜΕΝΑ ΥΛΙΚΑ.

COMPLAINTS: Complaints may be submitted initially written or vocal. Subsequently they need to be filed including the test performance and results in writing in case of analytical reasons.

WARRANTY: The product is warranted to be free from material defects within the specific shelf life and to comply with product specifications delivered with the product. The product must be used according to the Intended use, all instructions given in the instructions for use and within the product specific shelf life. Any modification of the test procedure or exchange or mixing of components of different lots could negatively affect the results. These cases invalidate any claim for replacement.

LIMITATION OF LIABILITY: IN ALL CIRCUMSTANCES THE EXTENT OF MANUFACTURER'S LIABILITY IS LIMITED TO THE PURCHASE PRICE OF THE KIT(S) IN QUESTION. IN NO EVENT SHALL MANUFACTURER BE LIABLE FOR ANY INCIDENTAL OR CONSEQUENTIAL DAMAGES, INCLUDING DAMAGES FOR LOST PROFITS, LOST SALES, INJURY TO PERSON OR PROPERTY OR ANY OTHER INCIDENTAL OR CONSEQUENTIAL LOSS.

	IBL International GmbH Flughafenstr. 52A, 22335 Hamburg, Germany	Tel.: + 49 (0) 40 532891 -0 Fax: -11 E-MAIL: IBL@IBL-International.com WEB: http://www.IBL-International.com
--	---	--