





## PRESTIGE 24i LQ CK

Cat. No 4-220, 4-420

(EN)

### INTENDED USE


Diagnostic kit for determination of creatine kinase activity used in automatic analysers Prestige 24i, Biolis 24i, Prestige 24i Premium, Biolis 24i Premium and Biolis 30i. The reagents must be used only for *in vitro* diagnostic, by suitably qualified laboratory personnel, only for the intended purpose, under appropriate laboratory conditions.

### INTRODUCTION

Creatine kinase (CK) catalyzes the transfer of phosphate group between creatine phosphate and adenosine diphosphate (ADP). The product of this reaction is adenosine triphosphate (ATP) – molecular source of energy. CK is a dimer, composed of two different subunits called M and B. Three different isoenzymes formed from these subunits are found in brain and smooth muscle (BB), skeletal muscle (MM) and cardiac muscle (MM and MB). Increased level of CK is usually the result of muscle injury, myocardial or pulmonary infarction.

### METHOD PRINCIPLE

Optimized kinetic method according to International Federation of Clinical Chemistry (IFCC).



The rate of absorbance changes at  $\lambda=340$  nm is directly proportional to creatine kinase activity.

### REAGENTS

#### Package

	Cat. No 4-220 (24-TRAY)	Cat. No 4-420 (36-TRAY)
1-Reagent	2 x 40 ml	2 x 22.5 ml
2-Reagent	2 x 10.5 ml	2 x 6 ml

The reagents when stored at 2-8°C are stable up to expiry date printed on the package. Stability on board of the analyser at 2-10°C: Prestige 24i - 12 weeks, Biolis 24i Premium – 12 weeks.

### Concentrations in the test

#### 1-Reagent

imidazole buffer	100 mmol/l
glucose	20 mmol/l
N-acetylcysteine	20 mmol/l
magnesium acetate	10 mmol/l
EDTA	2 mmol/l
NADP	2 mmol/l
ADP	2 mmol/l
AMP	5 mmol/l
HK	> 2.5 U/ml

### REFERENCE VALUES<sup>8</sup>

serum	37°C	
female	< 167 U/l	< 2.78 µkat/l
male	< 190 U/l	< 3.17 µkat/l

It is recommended for each laboratory to establish its own reference ranges for local population.

### QUALITY CONTROL

For internal quality control it is recommended to use, with each batch of samples, the CORMAY SERUM HN (Cat. No 5-172) and CORMAY SERUM HP (Cat. No 5-173).

For the calibration of automatic analysers systems the CORMAY MULTICALIBRATOR LEVEL 1 (Cat. No 5-174; 5-176) and LEVEL 2 (Cat. No 5-175; 5-177) are recommended.

The calibration curve should be prepared every 12 weeks (Prestige 24i, Biolis 24i Premium), with change of reagent lot number or as required e.g. quality control findings outside the specified range.

### PERFORMANCE CHARACTERISTICS

The following results have been obtained using automatic analyser Biolis 30i. Results may vary if a different instrument or a manual procedure is used.

#### ▪ LoB (Limit of Blank):

1.0 U/l (0.017 µkat/l)

#### ▪ LoD (Limit of Detection):

2.0 U/l (0.033 µkat/l)

#### ▪ LoQ (Limit of Quantitation):

8.5 U/l (0.14 µkat/l)

#### ▪ Linearity:

up to 2800 U/l (46.67 µkat/l).

For higher activity, dilute the sample with 0.9% NaCl and repeat the assay. Multiply the result by the dilution factor.

#### ▪ Specificity / Interferences

Haemoglobin up to 0.156 g/dl, bilirubin up to 20 mg/dl, ascorbate up to 62 mg/l and triglycerides up to 1000 mg/dl do not interfere with the test.

#### ▪ Precision

Repeatability (run to run) n = 20	Mean [U/l]	SD [U/l]	CV [%]
level 1	138.0	0.82	0.59
level 2	421.3	2.29	0.54
Reproducibility (day to day) n = 80	Mean [U/l]	SD [U/l]	CV [%]
level 1	135.3	6.48	4.8
level 2	434.4	5.26	1.2

#### ▪ Method comparison

A comparison between CK values determined at **Biolis 30i** (y) and at **ADVIA SIEMENS 1800** (x) using 55 serum samples gave following results:

$$y = 0.922 x + 2.4994 \text{ U/l};$$

$$R = 0.999 \quad (\text{R} - \text{correlation coefficient})$$

### WASTE MANAGEMENT

Please refer to local legal requirements.

### LITERATURE

- DGKC: J. Clin. Chem. Clin. Biochem.: 15, 249-254 (1977).
- The Committee on Enzymes of The Scandinavian Society for Clinical Chemistry and Clinical Phys.: Scand. J. Clin. Lab. Invest. 36, 1-5 (1979).
- Lott J.A., Stang J.M.: Clin. Chem. 26/9, 1241-1250 (1980).
- Commission Enzymologie, Comité de Standardisation, Société Française de Biologie Clinique: Ann. Biol. Clin. 40, 138-149 (1981).
- Tietz N.W., ed. Clinical Guide to Laboratory Tests, 3rd ed. Philadelphia, PA: WB Saunders, 806-6 (1995).
- Burtis C.A., Ashwood E.R., ed. Tietz Textbook of Clinical Chemistry, 3rd ed. Philadelphia, PA: Moss D. W., Henderson A. R., 652 (1999).
- Alan H.B. Wu: Tietz Clinical Guide to Laboratory Tests, 4th ed. WB Saunders, 634, (2006).
- Dembińska-Kiec A., Naskalski J.W.: Diagnostyka laboratoryjna z elementami biochemii klinicznej, Volumed, 786, (1998).

Date of issue: 04. 2021.







## PRESTIGE 24i LQ CK

### • Biolis 30i

Item no	19	Item name	CK	Specimen	SERUM	OPTICAL	
<b>Data information</b>				<b>Aspiration volume</b>			
UNITS	U/L		TYPE	Double			
DECIMALS	1			SAMPLE	REAGENT 1	REAGENT 2	
<b>Analysis</b>				VOL. ( $\mu$ L)	12	200	40
METHOD	RATE method		BOTTLE (ml)				
Main Wave Length	340 nm		FIRST DIL.				
Sub Wave Length	660 nm						
<b>CORRELATION (Y= AX + B)</b>				<b>Data processing read</b>			
A =	1			START	END		
B =	0		MAIN	37	53		
<b>Blank value</b>				SUB			
•WATER		° REAGENT	ABS LIMIT	-3	TO	3	
<b>Calibration</b>				<b>Collection value</b>			
TYPE	Linear 2		END POINT	2.5			
STABILITY			LINEARITY CHECK (%)	90			
				<b>Prozone check</b>			
MALE		FEMALE	START	END	LIMIT (%)		
LOW	HIGH	LOW	HIGH	FIRST			
0	190	0	167	SECOND			
				MINIMUM ABS.			
°HIGH			MEAN				
•LOW			VARIATE				

Item No	19	Item Name	CK	Specimen	SERUM	OPTICAL
<b>Reference intervals</b>				<b>Auto rerun</b>		
MALE		FEMALE		•ON	°OFF	
LOW	HIGH	LOW	HIGH			
0	190	0	167			
<b>Panic range</b>				<b>Auto rerun range (conc.)</b>		
MALE		FEMALE		Re	Value	Dil.
LOW	HIGH	LOW	HIGH		8.5	
				Re	Value	Dil.
					2800	
				<b>Auto rerun condition (abs.)</b>		
MALE		FEMALE		LOWER	°ON	•OFF
LOW	HIGH	LOW	HIGH	HIGH	°ON	•OFF
				<b>Auto rerun condition (prozone)</b>		
MALE		FEMALE		°ON	•OFF	
LOW	HIGH	LOW	HIGH			
				<b>Dilution</b>		
MALE		FEMALE		•DIL 1	° DIL 2	
<b>Reaction check</b>				<b>VH CHECK</b>		
°ON		•OFF		°ON	•OFF	
CHECK						
LOW						
HIGH						
<b>VL CHECK</b>		<b>VH CHECK</b>				
°ON	•OFF	°ON	•OFF			

Data wydania / Date of issue / Дата создания: 04.2021.