

Instruction for DS-Ref Ref. Filling Solution

**【Product Name】**

DS-Ref Ref. Filling Solution

**【Package Specification】**

10mL/bottle REF:42046001

**【Intended Use】**

Used as filling solution for reference electrode.

**【Test Principle】**

Not Applicable

**【Main Composition】**

Saturated KCl solution and Preservative.

**【Storage and Transportation】**

Stable for 2 years when stored at 2-30 °C in the shade. The products should be transported with outer package.

**【Applicable Instruments】**

Electrolyte Analyzers

**【Sample Requirements】**

Not Applicable

**【Test Methods】**

Fill Ref. filling solution into reference electrode. For more Instrument operations, please refer to electrolyte analyzer operation manual.

**【Reference Range】**

Not Applicable

**【Interpretation of Test Results】**

Not Applicable

**【Calibration and QC】**

Not Applicable

**【Limitations of Testing Methods】**

Not Applicable








**【Performance Indicator】**





Not Applicable

**【Precaution】**

1. Please seal it properly after using, to avoid long-time exposure in the air.
2. Avoid skin contact .Wear gloves when operating.
3. If the reagent is carelessly spattered onto the skin, eyes, etc., flush with clean water immediately, and if mistakenly swallowed, seek medical advice.
4. The best volume of solution for normal operation is 2/3 level of the intra-cavity.
5. Do not use the solution when package is damaged.

**【Icon Illustration】**

Label	Meaning
	Date of manufacture
	In vitro diagnostic medical device
	Manufacturer
	Batch code
	Temperature limit
	Use-by date
	Authorized representative in the European Community

	Catalogue number
	CE Marking
	Volume
	Main component

**【Training information】**

Please refer to the service manual.

**【Help information】**

If you need help please contact after-sales.

**【Trouble shooting】**

Please contact after-sales.

**【References】**

Not Applicable.

**【Manufacturer】**



Genrui Biotech Inc.

Address: 4-10F, Building 3, Geya Technology Park, Guangming District, 518106, Shenzhen, China

Web: www.genrui-bio.com

**【Medical Device Manufacturing Enterprise Permit No.】**

Guangdong SFDA (State Food and Drug Administration) authorized Medical Device Manufacturing Permit No. 20041046.

**【Guarantee and Technical Support】**

If invalid message repeats or need technical support, please contact Genrui Customer Service and Support Center.

**【Instruction Approved and Revised Date】**

Approved date: November 6<sup>th</sup> 2015

Revised date: October 19<sup>th</sup> 2018



Lotus NL B.V.  
Koningin Julianaplein 10, 1e Verd  
2595AA, The Hague, Netherlands  
Email : peter@lotusnl.com