

Pediatric Calibrator

Liquid human based calibrator for use as a calibrator in immunoturbidimetric assays

REF	Cont.	
A00705	1x 1 mL	Pediatric Calibrator
A00706	1x 5 mL	Pediatric Calibrator

COMPOSITION

The Pediatric Calibrator is a defibrinated human plasma, liquid stabilized and filtered through 0.2 µ.
 Sodium azide 0.095 %

PREPARATION

The Calibrator is liquid, ready to use.

STABILITY AND STORAGE

Storage: at 2 – 8 °C
 Stability: up to the expiration date
 Stability after opening: 6 weeks at 2 – 8 °C
 DO NOT FREEZE!

DIRECTIONS FOR USE

1. Use this Calibrator in the same manner as specified with the reagent and/or instrument being used.
2. Allow the Calibrator to equilibrate to room temperature (22 – 28 °C) prior to use.
3. The Calibrator should not be allowed to stand for repeated long periods of time (up to 6 hours) at room temperature. Keep vials tightly capped at all times when not in use to avoid microbial contamination. Keep vials refrigerated when not in use

WARNINGS AND PRECATIONS

1. For in Vitro Diagnostic use.
2. Each individual donation intended for use in manufacture of protein calibrator was tested for hepatitis B surface antigen (HBsAg), anti-hepatitis C (anti-HCV) and anti-HIV 1 and HIV 2 by FDA required tests. Since no test method can assure that products derived from human blood do not contain HIV-1/2 and Hepatitis B and Hepatitis C virus, this material and all patient samples should be handled as though capable of transmitting infectious diseases.
3. Reagents containing sodium azide must be handled with due caution: Do not ingest or allow to contact skin or mucous membranes! Sodium azide can form explosive azides when contacting heavy metals such as copper or lead.

WASTE MANAGEMENT

Please refer to local legal requirements.

LOT SPECIFIC ASSAY DATA

Values and expiry date are lot specific.

LOT: MPP009		EXP: 2016/01
Parameter	Assigned Value *	
α-2 Macroglobulin	30.90 mg/dL	
Transferrin	39.10 mg/dL	
IgA	29.80 mg/dL	
IgG	134.80 mg/dL	
IgM	17.90 mg/dL	
Kappa Light Chain **	35.90 mg/dL	
Lambda Light Chain **	18.90 mg/dL	
* Values traceable to ERM-DA470 from IFCC. ** Values based on Dade Behring Material.		

Pediatrischer Kalibrator

Flüssigkalibrator auf Humanbasis für quantitative Bestimmungen mittels immunturbidimetrischer Methode

REF	Cont.	
A00705	1x 1 mL	Pediatrischer Kalibrator
A00706	1x 5 mL	Pediatrischer Kalibrator

ZUSAMMENSETZUNG

Der Pediatrische Kalibrator basiert auf von Fibrin befreitem Humanplasma, flüssig-stabil und durch einen Filter mit 0,2 µ Porengröße gefiltert.
 Natriumazid 0.095 %

VORBEREITUNG

Der Kalibrator ist flüssig und gebrauchsfertig.

STABILITÄT UND LAGERUNG

Lagerung: bei 2 – 8 °C
 Haltbarkeit: bis zum Ablaufdatum
 Stabilität nach den Öffnen: 6 Wochen bei 2 – 8 °C
 NICHT EINFRIEREN!

TESTDURCHFÜHRUNG

1. Der Kalibrator muss wie im Beipacktext des verwendeten Reagenz/Instruments beschrieben verwendet werden.
2. Der Kalibrator muss vor Gebrauch auf Raumtemperatur (22 – 28 °C) gebracht werden.
3. Der Kalibrator sollte nicht für wiederholt längere Zeit (bis zu 6 Stunden) bei Raumtemperatur gelagert werden. Die Fläschchen immer fest verschlossen halten, wenn sie nicht in Gebrauch sind, um mikrobielle Kontamination zu vermeiden. Wenn nicht in Gebrauch, die Fläschchen gekühlt lagern.

WARNUNGEN UND VORSICHTSMASSNAHMEN

1. Für die In-Vitro-Diagnostik.
2. Jede Spende, die für die Verwendung in der Produktion von Proteinkalibratoren gedacht ist, wurde auf Hepatitis B Oberflächenantigen (HBsAg), Anti-Hepatitis C (Anti-HCV) und Anti-HIV 1 und HIV 2 unter Verwendung einer FDA-geprüften Methode getestet. Da keine Methode 100%ige Sicherheit bietet, dass Produkte humanen Ursprungs kein HIV-1/2, Hepatitis B oder Hepatitis C Viren enthalten, sollte dieses Material und sämtliche Patientenproben wie potentiell infektiöses Material gehandhabt werden.
3. Reagenzien, die Natriumazid beinhalten, müssen mit Vorsicht behandelt werden: Nicht verschlucken und Kontakt mit Haut und Schleimhäuten vermeiden! Natriumazid kann explosive Azide bilden, wenn es mit Schwermetallen wie Kupfer oder Blei in Kontakt kommt.

ABFALLBESEITIGUNG

Die lokalen Bestimmungen sind zu beachten.

LOTSPEZIFISCHE WERTE

Werte und Ablaufdatum sind lotspezifisch.

LOT: MPP009		EXP: 2016/01
Parameter	Bestimmter Wert *	
α-2 Macroglobulin	30.90 mg/dL	
Transferrin	39.10 mg/dL	
IgA	29.80 mg/dL	
IgG	134.80 mg/dL	
IgM	17.90 mg/dL	
Kappa Leichtkette **	35.90 mg/dL	
Lambda Leichtkette **	18.90 mg/dL	
* Werte auf ERM-DA470 des IFCC rückführbar. ** Werte basieren auf Dade Behring Material.		

